

# Cardiosave IABP Operation - Transport





## Quick Reference Guide/Cardiosave IABP software C.06

(For use outside the US only)




**Note:** This is an abbreviated guide. For complete instructions, refer to IAB Instructions for Use and Cardiosave Operating Instructions.

## Intra-aortic Balloon Sizing Guide

 <p>&lt; 5'0" ( &lt; 152 cm)</p>	 <p>5'0" to 5'4" (152 cm - 162 cm)</p>	 <p>5'4" to 6'0" (162 cm - 183 cm)</p>	 <p>5'4" &amp; taller (≥ 162 cm)</p> <p>6' 5'</p>
<p>30 cc</p>	<p>40 cc</p>	<p>50 cc</p>	
<p>Mega 7.5 Fr.</p>	<p>Sensation Plus 7.5 Fr. Mega 7.5 Fr.</p>	<p>Sensation Plus 8 Fr. Mega 8 Fr.</p>	
<p>25 cc</p>	<p>34 cc</p>	<p>40 cc</p>	
<p>Linear 7.5 Fr.</p>	<p>Sensation 7 Fr. Linear 7.5 Fr.</p>	<p>Sensation 7 Fr. Linear 7.5 Fr.</p>	

**Note:** This information is to be used as a guidance only. Clinical information and patient factors such as torso length should be considered when selecting the appropriate balloon size. Sensation and Sensation Plus are fiber-optic IAB catheters.

# Cardiosave Rescue initial Set-up



## Initial set-up using a Getinge Fiber-optic IAB (continued)



Verify  
AUTO

Press  
START



Verify Aug. Alarm is set 10 mmHg below patient's augmented diastolic pressure (approximately 3 minutes after initiation of assist).

### Pressing the START key

- Automatically purges and fills IAB
- Automatically performs an *in vivo* calibration
- Automatically selects most appropriate lead and trigger
- Automatically sets inflation and deflation timing

**Note:** With a Getinge fiber-optic IAB, there is no need to zero. Calibration occurs automatically after pressing START. Operator may invoke a calibration anytime by pressing and holding CALIBRATE PRESSURE key for 2 seconds, while assisting.



Check fill status of the helium tank. A fully charged internal helium reservoir will provide approximately 36 fill cycles.

## Initial Set-up using a Conventional IAB (continued)



Verify  
**AUTO**

Press  
**START**



Verify Aug. Alarm is set 10 mmHg below patient's augmented diastolic pressure (approximately 3 minutes after initiation of assist).

### Pressing the **START** key

- Automatically purges and fills IAB
- Automatically selects most appropriate lead and trigger
- Automatically sets inflation and deflation timing


## Portable Operation

### Getinge recommends:


- Sufficient supply of fully charged batteries for use during transport.
- Use of the Transport Power Supply for AC operation during transport.
- Verifying internal helium reservoir is full when using the Transport System.
- System must be properly secured in the transport vehicle.




## Removing Pump Console from the Cart




1 Release latch located below pump console (ensure wheels are locked).




2 Grab handle and slowly slide console out.  
**Note:** 3 audio tones will sound.



3 Grab handles located on top and front of console, then remove from Hospital Cart.



4a




4b

4a Push button to release pop-up mount.  
4b Pull UP pop-up mount to lock in place.




5

5 Squeeze latches located below Monitor and lift to remove from Hospital Cart.



6

6 Squeeze latches and attach to pop-up mount, then release latches. Ensure Monitor is securely attached.



7

7 Squeeze latch below handle and lift straight up until wheels extend outward and handle locks into extended position.




8

8 Tilt Transport System on wheels and begin transport.




Rescue icon will be displayed on Touchscreen.

## Inserting the Pump Console into the Hospital Cart





Squeeze latch below handle. Push straight down until wheels retract and handle is fully collapsed.



Squeeze latches located below Monitor and lift to remove from Pump Console.




Squeeze latches and attach to display mount, then release the latches. Ensure monitor is securely attached.




**4a** Push and hold button on console to unlock monitor mount.

**4b** Push down to lock into place, then release button.



Grab handles located on top and front of console, then lift into Hospital Cart.




Grab handle and slowly slide console into Hospital Cart until it locks into place.

**Note:** An audible click will be heard when console is locking into cart and 3 audio tones will sound.



Ensure Hybrid icon (7a) is displayed on touchscreen when console has been successfully installed into cart. Plug Hospital Cart power cord into a compatible grounded AC receptacle. Confirm AC operation by presence of AC Plug icon (7b). If Hybrid icon/AC Plug icon are not present after installing console into the cart:

- Release latch on Hospital Cart located below Pump Console
- Grab handle and slowly slide Console out approximately one quarter of the way
- Repeat steps 6-7a-b to ensure Pump Console was successfully installed into Hospital Cart





## Helium Supply

Ensure the internal helium reservoir is full before using **Cardiosave** for transport.

If the helium icon is not showing full, the following actions should be taken:


### Cardiosave Hybrid:

- Return the pump console to the hospital cart for a minimum of 30 seconds.
- Ensure the helium icon indicates full.
- Remove pump console from cart and continue with transport (refer to page 6).

### Cardiosave Rescue:

- Attach the Helium Refilling Station for a minimum of 30 seconds.
- Ensure the helium icon indicates full.
- Disconnect Helium Refilling Station and continue with transport.

**Caution:** The Helium Refilling Station is not intended for use in transport. It is intended to be used in office buildings, aircraft hangars, or similar environments, and should not be within the vicinity of a patient.



Helium tank icon displays a full tank.



Helium Refilling Station

## Helium use from Internal Reservoir

The internal helium reservoir contains sufficient helium to provide approximately 36 fill cycles at full capacity. With every Autofill, helium will be depleted in the following approximate amounts:


Autofill Condition	Helium Used
Pump is powered off, powered on, and IAB fill performed to restart therapy	6 Autofill cycles
IAB disconnected and reconnected and IAB fill performed to restart therapy	6 Autofill cycles
Autofill performed due to a Gas Gain in IAB Circuit alarm, IAB Disconnected alarm, or an Autofill Failure alarm	6 Autofill cycles
Autofill every 2 hours	1 Autofill cycle
Autofill every 1000 feet (305 meters) of altitude increase during ascent	1 Autofill cycle
Autofill every 2000 feet (610 meters) of altitude decrease during descent	1 Autofill cycle

**Note:** The supply of helium in the internal helium reservoir will deplete more rapidly when an autofill is performed when the system is powered on, when the catheter is disconnected and reconnected or when an autofill is performed due to a gas loss, catheter disconnect, or autofill failure alarm.


### Effects of altitude changes during air transportation

For proper operation during air transport, IABP balloon pressure must adapt to local atmospheric pressure. The system will automatically purge and fill the IAB when local atmospheric pressure decreases by 25 mmHg or increases by 50 mmHg. These pressure changes occur approximately every 1,000 feet (305 meters) of increase in altitude or 2000 feet (610 meters) of decrease in altitude.


## Viewing Battery Status on Monitor Display (examples)




Plugged into AC power outlet and batteries are fully charged.




Lit green circle indicates battery 1 is in use. Battery 2 is fully charged and available for use when battery 1 is depleted.




Battery 1 is depleted, thus battery 2 is currently being used.




Battery 2 has less than 30 minutes of charge remaining (Low Battery message is displayed).




Battery 2 has less than 5 minutes of charge remaining.



IABP has been plugged into AC power outlet and battery 2 is being charged.




Transport Power Supply installed in Battery Bay #2 is plugged into AC receptacle and is in use. The battery in Battery Bay #1 is not being charged.




Empty battery bay. No backup battery or Transport Power Supply detected in battery bay.

## Viewing Battery Status on Battery



Battery is approximately 100% charged.


**Note:** Each LED represents a charge of approximately 20%.  
Battery run time: approximately 90 minutes each.




Battery installed in  
Battery Bay 1.

Transport Power Supply  
installed in Battery Bay 2.


## Changing the Battery



Turn knob to remove  
battery from Battery Bay.



Slide battery OUT.



Slide charged battery IN.




While holding battery in  
bay, turn knob to lock  
battery in place.




Use care to avoid  
dropping the battery.



## IAB Transport Fixation with the Mounting Plate



1 Center Cardiosave on open end of Transport Mounting Plate. Ensure pump's wheels are fully retracted before rolling into position on the plate.



2 Roll Cardiosave into Transport Mounting Plate between guide tracks until seated on guide pins and firmly braced against rear track.



3a Lift swing bolt into upright position.

3b Turn bolt clockwise until tight. Ensure Cardiosave is firmly locked in place.

## Technical Specifications

### Weight\*

	kg	lbs
Monitor	3.6	8
Battery Pack (0146-00-0097)	1.4	3
Pump Console	17.7	39
<b>Transport Configuration</b> (including monitor and pump console with 2 batteries)	24.1	53

\* All weights  $\pm$  5%

### Dimensions

Transport and Display	
Display Closed	57.2 cm H x 40.6 cm D x 33.0 cm W 22.5" H x 16" D x 13" W
Display Open 90°	78.0 cm H 30.7" H

### Operating Ambient

Operating Temperature	10° C to 40° C (50° F to 104° F)
Operating Humidity	15% to 85% Relative Humidity (non-condensing)
Operating Altitude	-1250 feet to 12,000 feet (795 mmHg to 483 mmHg) (1060 hPa to 644 hPa)


# Cardiosave Operation Modes

## Auto Operation Mode

- Automatic lead and trigger selection
- Automatic and continuous inflation and deflation timing management
  - User has ability to fine-tune deflation timing
- Automatic management of irregular rhythms
- Automatic *in vivo* calibration (when using a Getinge fiber-optic IAB)

## Semi-Auto Operation Mode


- Operator selects most appropriate lead and trigger source
- Operator establishes timing, then Cardiosave automatically adjusts timing with heart rate and rhythm changes
- Automatic management of irregular rhythms
- Automatic *in vivo* calibration (when using a Getinge fiber-optic IAB)



## Cardiosave Monitor Display and Touchscreen




Monitor Display



Touchscreen




# Cardiosave Lock Screen Feature



Touchscreen Unlocked

Touchscreen will **Lock**:

- Automatically after 2 minutes of inactivity
- When operator presses LOCK SCREEN key for 2 seconds



Touchscreen Locked

Touchscreen will **Unlock**:

- Automatically with any Technical, High, Medium, or Low Priority Alarm
- When operator presses UNLOCK SCREEN key



# Cardiosave Triggers

## Triggering

- A Trigger is the signal that Cardiosave uses to identify the beginning of the next cardiac cycle
- When Cardiosave recognizes the trigger event, it will deflate the balloon if not already deflated
- Trigger Source keys are only active while in Semi-Auto operation mode




## Cardiosave Triggers

### ECG

#### Trigger event is the R-Wave

- Trigger of choice when an adequate R-Wave is present
- Pacer spikes are automatically rejected



### Pressure

#### Trigger event is the systolic upstroke

- Trigger of choice (with a regular rhythm) when an adequate R-Wave is not present
- A fixed pressure threshold can be manually set while in Semi-Auto operation mode



### Pacer V/AV

#### Trigger event is the Ventricular pacer spike

- Typically used when ECG triggering is unsuccessful and a V or AV pacer is being used
- Must be 100% paced
- Only available in Semi-Auto operation mode

### Pacer A

#### Trigger event is the R-Wave

- Recommended only if atrial pacer tails are interfering with R-Wave detection while in ECG trigger
- Only available in Semi-Auto operation mode

### Internal

#### Trigger event is asynchronous at a fixed rate of 80 BPM

- Only used when there is no mechanical cardiac cycle (i.e.: cardiopulmonary bypass or asystole)
- Rate can be adjusted from 40 to 120 BPM
- Only available in Semi-Auto operation mode

# Theory of Counterpulsation Therapy



**Inflation: increases supply of oxygen to the myocardium.**

## How it works

- Balloon inflates at onset of diastole (when aortic valve closes)
- Displaces blood, causing an increase in aortic pressure

## Benefits

- Increases coronary artery perfusion
- Increases mean arterial pressure



**Deflation: decreases demand for oxygen by the left ventricle.**

## How it works

- Balloon deflates just prior to systolic ejection (before aortic valve opens)
- Results in a rapid decrease in aortic pressure

## Benefits

- Decreases afterload
- Decreases cardiac workload
- Increases cardiac output

## Timing

Timing refers to the positioning of inflate and deflate points on the arterial pressure waveform.


### Proper IABP Timing

#### Inflation


- Occurs at the dicrotic notch
- Appears as a sharp “V”
- Ideally diastolic augmentation rises above systole

#### Deflation

- Occurs just prior to systolic ejection
- Results in a reduction in assisted end diastolic pressure
- Results in a reduction in assisted systolic pressure



## Timing Errors



### Early Inflation


Inflation of IAB prior to aortic valve closure.

#### Waveform characteristics

- Inflation of IAB prior to dicotic notch
- Diastolic augmentation encroaches onto systole (may be unable to distinguish)

#### Physiologic Effects

- Potential premature closure of aortic valve
- Potential increase in LVEDV/LVEDP/PCWP
- Increased left ventricular wall stress or afterload
- Aortic regurgitation
- Increased  $MVO_2$  demand



### Late Inflation

Inflation of IAB markedly after closure of aortic valve.


#### Waveform characteristics

- Inflation of IAB after dicotic notch
- Absence of sharp "V"
- Sub-optimal diastolic augmentation

#### Physiologic Effects

- Sub-optimal coronary artery perfusion

## Timing Errors



### Early Deflation


Premature deflation of IAB during diastolic phase.

#### Waveform characteristics

- Deflation of IAB is seen as a sharp drop following diastolic augmentation
- Sub-optimal diastolic augmentation
- Assisted end diastolic pressure may be equal to or less than unassisted end diastolic pressure
- Assisted systolic pressure may rise

#### Physiologic Effects

- Sub-optimal coronary perfusion
- Potential for retrograde coronary and carotid blood flow
- Angina may occur as a result of retrograde coronary blood flow
- Sub-optimal afterload reduction
- Increased  $MVO_2$  demand



### Late Deflation

Deflation of IAB after aortic valve has opened.

#### Waveform characteristics



- Assisted end diastolic pressure may be equal to or higher than unassisted end diastolic pressure
- Rate of rise of assisted systole is prolonged
- Diastolic augmentation may appear widened

#### Physiologic Effects


- Afterload reduction is essentially absent
- Increased  $MVO_2$  consumption due to left ventricle ejecting against a greater resistance and a prolonged isovolumetric contraction phase
- IAB may impede left ventricular ejection and increase afterload

## Proper Care of Inner Lumen

- Minimize length of pressure tubing
- Use **only** low compliance pressure tubing
- Elevate flush bag at least 3' (91.44 cm) above transducer
- A 3 cc/hour continuous flow through inner lumen is recommended
- If inner lumen becomes damped
  - Aspirate and discard 3 cc of blood
  - If **unable** to aspirate blood, consider inner lumen clotted, cap lumen, provide alternate pressure source
  - If **able** to aspirate blood, fast flush to clear pressure tubing for **at least 15 seconds** (with IABP on Standby)
- **Do not** sample blood from inner lumen
  
- The saline pole is included with the IABP as a convenient location for the flush bag, when needed
- Never place fluids on top of the IABP and **DO NOT** hang flush bag and/or tubing directly over the IABP
- In case of accidental spillage, wipe clean immediately and have the unit serviced



## Patient Assessment



### Assessment

### Corrective Action

#### Radial pulses

Left radial pulse weak or left arm ischemia.

Check position of IAB.

#### Insertion site

Excessive bleeding from insertion site.

Apply pressure, ensure distal flow.

#### Pedal pulses

Limb ischemia detected.

Consider removing IAB, consider insertion via opposite limb.

#### IAB inner lumen flush line

Pressure waveform damped (if using a conventional IAB).

Aspirate inner lumen. If line patent, flush for 15 seconds (with IABP on Standby).

#### Urine output

Urine output low.

Check position of IAB.

#### IAB catheter tubing

Blood observed in catheter tubing.

STOP pumping and prepare for IAB removal.




## Suspected IAB perforation

### Evidence of IAB perforation

- Blood or fluid may be seen in extracorporeal tubing or catheter extender as evidenced by:
  - Bright red blood
  - Dried blood particles
  - Serosanguineous fluid
- Sudden change in diastolic augmentation pressure
- Potential IAB pump alarms

### Potential IABP Alarms

Periodically check IAB catheter tubing for blood both throughout therapy and when the below alarms occur. Do not bypass these alarms, and please pay close attention to alarm notifications listed below, as these alarms may help identify a perforated balloon earlier, preventing any blood from traveling into the IABP.



#### During an autofill

##### Potential alarms

- Autofill Failure – Blood Suspected
- Autofill Failure

#### Outside an autofill

##### Potential alarms


- Gas Loss in IAB Circuit
- Gas Gain in IAB Circuit
- IAB Catheter Restriction

## Suspected IAB perforation

### Blood detected management

If any blood is noted or perforation is suspected, the following procedure must be performed immediately.

#### 1 Standby




Stop pumping by placing IABP console in Standby

#### 2 Disconnect



Disconnect catheter extender tubing from IABP console to allow balloon to deflate

#### 3 Clamp



Clamp extracorporeal tubing between white y-fitting and male connector

### Clinical Considerations:

- Place patient in Trendelenburg as tolerated to guide any residual helium to travel away from the head vessels.
- Notify physician, and prepare for IAB catheter removal.
- Consider IAB catheter replacement, if patient's condition warrants.
- If blood is suspected of having entered the pump, take pump out of service. It should be evaluated before use in another patient by Biomed/Technical Service to determine if replacement of contaminated components are necessary.

# Alarms

## Augmentation Below Limit Set



### Probable Cause

Hemodynamic status has changed:  $\uparrow$ HR,  $\downarrow$ SV,  $\downarrow$ MAP.



Alarm limit set too high.

### Corrective Action

Attempt to optimize patient's hemodynamic status.

Press AUG. ALARM key, decrease limit.

## Autofill Failure



### Probable Cause

IAB disconnected.

Helium tank is closed.

Helium tank is empty.

Incorrect IAB catheter extender tubing length.

### Corrective Action

Attach IAB catheter.


Open helium tank.

Change helium tank.

Ensure only one IAB catheter extender tubing is connected from IAB to pump.

## Alarms

### IAB Catheter Restriction



#### Probable Cause

Restriction in IAB catheter or tubing.

#### Corrective Action

Relieve restriction if possible, press START.




Membrane has not completely unfolded.

IAB remains in sheath.

Manually inflate and deflate IAB.

Check markings on IAB and if IAB has not exited sheath, refer to IFU to reposition sheath relative to IAB catheter.

### IAB Disconnected



#### Probable Cause


IAB catheter or extender tubing is disconnected.

#### Corrective Action

Reattach IAB, press START.

# Alarms

## Prolonged Time in Standby






### Probable Cause

IABP has been in STANDBY for at least 10 minutes.

### Corrective Action

Verify whether it is appropriate to resume pumping.

## Gas Loss in IAB Circuit



### Probable Cause

A helium loss has been detected in IAB circuit.

### Corrective Action

If blood observed - STOP pumping. Prepare for removal of IAB.

If blood is not observed, verify connections are tight.

If appropriate, perform an Autofill, then press START to resume pumping.

**Note:** The IAB Catheter should not remain inactive (not inflating and deflating) for more than 30 minutes because of the potential for thrombus formation.

To keep IAB catheter active in the event of pump failure, manually inflate and deflate the IAB as follows:


1. Detach catheter extender from IAB catheter.
2. Attach 3-way stopcock and syringe to IAB catheter's male luer fitting.
3. Aspirate to ensure blood is not returned from catheter.
4. Inflate IAB with 40 cc air or helium and immediately aspirate. Repeat every 5 minutes while catheter is inactive.

**Warning:** Never inject air into inner lumen (female luer hub).

## Arterial Pressure Surveillance Alarm

### Cardiosave 3rd edition Compliant\* / Cardiosave with select software upgrade

- System performs Flat Arterial Pressure (A.P.) Surveillance checks and monitors for sustained loss of pulsatility on A.P. trace when pulsatility is expected
- If a sustained loss of pulsatility is detected, NO PRESSURE SOURCE AVAILABLE Medium Priority Alarm will occur
- Pulsatility surveillance can be suspended by pressing DISABLE PULSATILE TEST key in the Sources menu when a well understood situation creates a nuisance alarm condition
- When zeroing a pressure transducer, press the Zero Pressure key within 7 seconds after venting the transducer to avoid a NO PRESSURE SOURCE AVAILABLE Alarm



#### No Pressure Source Available

##### Probable Cause

No DIRECT or EXTERNAL arterial pressure source connected.

##### Corrective Action

Ensure fiber-optic sensor cable is connected.

If transducer in use, ensure pressure cable is connected to transducer and IABP.

If A.P. source unavailable from catheter, provide an A.P. signal from external monitor to IABP using interface cable.

\*Cardiosave 3rd edition compliant systems have a serial number (located on the back of the system) that begins with the letters CB.

## Arterial Pressure Surveillance Alarm, (continued)

### No Pressure Source Available

Probable Cause	Corrective Action
A.P. transducer vented to atmosphere for more than 10 seconds.	Check stopcocks to ensure transducer is closed to atmosphere.
Pressure monitoring tubing has become disconnected.	Verify monitoring tubing is securely connected.
Pressure monitoring lumen may be clotted.	Attempt to aspirate. If unable to aspirate, discontinue use of arterial lumen and provide an alternate A.P. source. If alternate A.P. source is not available, override the alarm by pressing the DISABLE PULSATILE TEST in the Sources Menu.
Defective pressure transducer or transducer cable.	Replace transducer or transducer cable. If alarm persists, provide an alternate A.P. source. If an alternate A.P. source is not available, override the alarm by pressing the DISABLE PULSATILE TEST key in the Sources menu.

## Alarms

### Unable to Update Timing

#### Probable Cause

#### Corrective Action

Poor waveform quality.

Check cable connections.  
Verify transducer was not left vented, if in use.

If transducer is in use, aspirate and flush arterial pressure line.

If problem persists, switch operation mode to SEMI-AUTO, verify TRIGGER SOURCE, adjust timing, resume pumping.

Sustained heart rate is less than 30 BPM or greater than 150 BPM.

Switch to SEMI-AUTO, verify TRIGGER SOURCE, adjust timing, resume pumping.

Poor diastolic augmentation.

If diastolic augmentation is poor, when AUGMENTATION level is set to MAX, attempt to improve patient's hemodynamic status.



# Alarms

## Unable to Calibrate Fiber-optic Sensor

### Probable Cause

Patient's pulse pressure is inadequate for calibration.

### Corrective Action

When patient's pulse pressure improves, press CALIBRATE PRESSURE key for 2 seconds while IABP is assisting.

Provide alternate A.P. source (i.e.: radial).

Extender tubing or balloon catheter may be restricted.

Relieve restriction.

Attempt calibration by pressing CALIBRATE PRESSURE key for 2 seconds while IABP is assisting.

## Fiber-Optic Sensor Calibration Postponed

### Probable Cause

A non-scheduled calibration update has been intentionally postponed because either patient's mean arterial pressure may be too low to pause assist or less than 15 minutes have elapsed since last calibration.

Pump is in STANDBY.

### Corrective Action

Assess patient to determine if a brief pause in assist would be tolerated, and if so, press CALIBRATE PRESSURE key for 2 seconds while IABP is assisting.

Provide alternate A.P. source (i.e.: radial).

Resume pumping, then press CALIBRATE PRESSURE key for 2 seconds to initiate a calibration.

## Alarms

### Fiber-optic Sensor Module Failure

Probable Cause	Corrective Action
There has been a failure of the internal Fiber-Optic Sensor Module in the IABP.	If a fiber-optic IAB is NOT in use, continue normal IABP use.
	If a Getinge fiber-optic IAB is in use, replace IABP with another Getinge IABP that supports the fiber-optic IAB.
	If replacement IABP is not available, provide alternate A.P. source (i.e.: radial).
	Contact Maquet Service for Fiber-Optic Sensor Module repair.

### Fiber-optic Sensor Failure

Probable Cause	Corrective Action
There is a failure in communication of the fiber-optic sensor signal with the IABP.	Unplug Fiber-Optic Sensor Connector and reconnect.
	Relieve any visible kinks in orange fiber-optic cable.
	If problem persists, disconnect Fiber-Optic Sensor Connector and provide alternate A.P. source (i.e.: radial).

## Connecting an Arrow IAB/IABP to a Getinge IABP

### Transferring Facility

- This patient will have an Arrow IAB connected to an Arrow IABP
- Before leaving facility, locate IAB catheter extender tubing supplied in Arrow IAB box, which connects an Arrow IAB to a Getinge IABP
- Take this IAB catheter extender tubing on transport with patient for use when arriving at receiving facility

### Receiving Facility

- When arriving at receiving facility, remove current IAB catheter extender tubing that connects an Arrow IAB to an Arrow IABP
- Connect appropriate end of IAB catheter extender tubing (that was brought from transferring facility) to Arrow IAB, then connect male luer fitting of IAB catheter extender tubing to back of Getinge IABP
- Set-up Getinge IABP per abbreviated instructions on page 3 of this Quick Reference Guide

## Connecting a Getinge IAB/IABP to an Arrow IABP

























### Transferring Facility

- This patient will have a Getinge IAB connected to a Getinge IABP
- Before leaving facility, locate Arrow Pump Adapter (APA) that connects a Getinge IAB to an Arrow IABP (may be supplied in Getinge IAB box or separately)
- Take the APA on transport with patient, for use when arriving at receiving facility

### Receiving Facility

- When arriving at receiving facility, place Getinge IABP on Standby and disconnect IAB catheter extender tubing from back of IABP
- Connect Arrow Pump Adapter (APA) to male luer fitting of Getinge IAB catheter extender tubing and connect to Arrow IABP
- Adjust volume setting on Arrow IABP, according to Operating Instructions, to match IAB catheter volume

## Cardiosave Symbols

Icon	Description	Icon	Description	Icon	Description
	Attention, refer to Operating Instructions		IABP		Technical Alarm
	Warning of Potential Injury or Health Risk		Do not place fluids on top of unit		High Priority Alarm
	Alternating Current (AC)		Patient		Medium Priority Alarm
	Battery		Patient Monitor		Low Priority Alarm
	Helium Tank		Vent		Audio Paused
	ECG		Fiber-optic cable		Alarm inhibited (off)
	Pressure		Trainer		Alarm inhibited (paused)
	Fiber-optic Connection Indicator		On/Off		Audio Alarms Off



**Notes**

A series of horizontal dotted lines for taking notes, spanning the width of the page.



**This page is intentionally left blank**



PN: 0002-08-9772 Rev C · Getinge, **GETINGE \***, and Maquet are trademarks or registered trademarks of Getinge AB, its subsidiaries, or affiliates in the United States or other countries · StatLock is a registered trademark of C.R. Bard, Inc. · Copyright 2022 Datascope Corp. · All rights reserved.

△ Refer to Instructions for Use for current indications, warnings, contraindications and precautions · 12/2022

**Getinge** · 45 Barbour Pond Drive, Wayne, NJ 07470 · USA

 **Datascope Corp.** · 1300 MacArthur Blvd., Mahwah, NJ 07430 · USA

[www.getinge.com](http://www.getinge.com)